

**SBE****SEA-BIRD ELECTRONICS, INC**

1808 - 136th Place NE, Bellevue, Washington 98005 USA

Telephone: (206) 643-9866 FAX: (206) 643-9954 Telex: 292915 SBEI UR

**Service Report**SBE Job Number: 18914RDate: 17 July 1998Customer: EMS**Customer Identified Problem:**

1. Calibrate/repair SBE 13 DO sensor, S/N 130338.

**Problems Found:**

1. SBE 13 DO sensor, S/N 130338.  
The DO module had reached the end of its life expectancy. Also, the "DO Module Housing" was found cracked.

**Repairs Performed:**

1. Calibrations and repairs performed on SBE 13 DO sensor, S/N 130338.  
Replaced the "DO Module Housing".  
Replaced DO module.  
Recalibrated the sensor.